



## **Calibration Services Purchase Order Criteria**

Calibration vendors shall meet the following requirements:

- Maintain a current ISO/IEC 17025 accreditation through an ILAC Full Member Body for the scope of the calibration service being requested or be subject to an assessment to verify compliance with ISO/IEC 17025 for the service provided.
- Provide information on the calibration certificate/report detailed in ISO/17025:2005 Clauses 5.10.2 and 5.10.4.
- Provide statement of uncertainties, including:
  1. Individual uncertainty values for each calibration measurement, or
  2. Total calculated measurement uncertainty and a statement that individual results are available upon request, or
  3. Statement of conformance to a tolerance (e.g. the uncertainty in the measurement system did not exceed 25% (4.1 test uncertainty ratio) of the acceptable tolerance for each characteristic calibrated, unless otherwise noted in the report) and a statement that individual results are available upon request.